

# Product Specifications

CL03.0213

HYDREX

## Buffer pH 4.00

Tampão pH 4.00  
Puffer pH 4.00  
Buffer pH 4.00  
Tampón pH 4.00  
Tampon pH 4.00  
Tampone pH 4.00

**For laboratory use, pH measurements, traceable to SRM from NIST  
pH 4.00 ( $\pm 0.02/20^{\circ}\text{C}$ ) Citrate-NaCl-NaOH**  
Traceability: Reference standards acc. NIST SRM 189c, 188, 185i,  
186lg, 186llg, 187e

Mol weight:  
Density: 1,01 g/ml  
HS Nr: 38229000  
CAS Nr:  
UN:  
ADR:

### Specifications:

pH	= 4.03 @ 0°C
pH	= 4.02 @ 10°C
pH	= 4.00 @ 20°C
pH	= 4.00 @ 25°C
pH	= 4.00 @ 30°C
pH	= 4.00 @ 40°C
pH	= 4.00 @ 50°C
pH	= 4.00 @ 60°C
pH	= 4.00 @ 70°C
pH	= 4.00 @ 80°C
pH	= 4.00 @ 90°C